



CERTIFICATE OF ANALYSIS

Artificial Urine for Corrosion Testing, Stabilized 1700-0016

Product: Artificial Urine for corrosion testing of urological implants

Part Number: 1700-0016

Lot Number: 906034

Production Date: 20 Jun 2019

Expiration Date: 20 Jun 2020

Primary Application: Product Testing

Acceptance Criteria:

1. pH = 6.0 ± 0.1 @ 25 ± 0.4 °C (Y/N) y

Storage Conditions: Dry at 8 to 25° C

Analysis Certified by: Jim Murphy **Date:** June 26, 2019
President
Title

Rev. A, 03/26/14