



CERTIFICATE OF ANALYSIS

Artificial Urine DIN EN 1616:1999 1700-0017

Product: Artificial Urine prepared according to DIN EN 1616:1999 standard procedure

Part Number: 1700-0017

Lot Number: 2104047

Production Date: 23 April 2021

Expiration Date: 23 April 2022

Primary Application: For testing of urethral catheters.

Acceptance Criteria:

1. pH = 6.6 ± 0.1 @ 25 ± 0.4 °C (Y/N)

Y

Storage Conditions: Keep frozen

Analysis Certified by: Jim Murphy
President

Title

Date: 23 April 2021

Rev. B, 05/08/19