



## CERTIFICATE OF ANALYSIS

### Artificial Urine DIN EN 1616:1999 1700-0017

**Product:** Artificial Urine prepared according to DIN EN 1616:1999 standard procedure

**Part Number:** 1700-0017

**Lot Number:** 508024

**Production Date:** 14 Aug 2015

**Expiration Date:** 14 Aug 2016

**Primary Application:** For testing of urethral catheters.

**Acceptance Criteria:**

1. pH =  $6.6 \pm 0.1$  @  $25 \pm 0.4$  °C (Y/N)

y

**Storage Conditions:** Keep refrigerated

**Analysis Certified by:** Jim Murphy  
President  
Title

**Date:** August 17, 2015

Rev. A, 03/26/14