



CERTIFICATE OF ANALYSIS

Carbamate Test Mixture

1700-0063

Product: Carbamate Test Mixture

Part Number: 1700-0063

Lot Number: 2101038

Production Date: February 10, 2021

Expiration Date: February 10, 2024

Primary Application: Carbamate Test Mixture is a solution of 12 compounds in methanol. Concentration of each compound is 2.5 µg/mL. This Test Mixture is intended to be used for establishing an elution profile and for troubleshooting. This product is not intended for calibration.

Acceptance Criteria:

1. The concentrations of Carbamate pesticides and Naphthol are within +/- 10% (Y/N)

Y

Storage Conditions: Keep frozen

Analysis Certified by: Jim Murphy **Date:** February 19, 2021
President
Title