



CERTIFICATE OF ANALYSIS

ChlorAC

1700-0132

Product: ChorAC

Part Number: 1700-0132

Lot Number: 009002

Test Start Date: 02 Sep 2010

Primary Application: Post-column analysis of carbamate pesticides

Acceptance Criteria:

1. pH = 3.0 ± 0.2 @ 25 ± 0.4 ° C (for a 3% solution of ChlorAC).
2. No interfering peaks in blank.
3. There are 12 peaks with qualitative Carbamate Test Standard (1700-0063).

Calibration Standard: Carbamate Test Standard, **PN** 1700-0063, **LN** 908045

Analysis Certified by:  **Date:** 17 Sep 2010

VP Ops
Title

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE