



# CERTIFICATE OF ANALYSIS

ChlorAC

1700-0132

Product: ChorAC

Part Number: 1700-0132

Lot Number: 103021

Production Date: 11 MAR 2011

Expiration Date: 11 MAR 2012


**Primary Application:** Preservative for waters samples in Post-column analysis of carbamate pesticides

**Acceptance Criteria:**

- 1. pH =  $3.0 \pm 0.2$  @  $25 \text{ }^\circ\text{C} \pm 0.4 \text{ }^\circ\text{C}$  (Y/N)       y
- 2. No interfering peaks in blank (Y/N)       y
- 3. 12 appropriate peaks in the chromatogram (Y/N)       y

**Sample:** Carbamate Test Mixture, **PN:** 1700-0063

**Storage Conditions:** Dry at 8°C to 25°C

Analysis Certified by:  Date: 21 MAR 2011  
VP OPR  
Title

**ACCEPTED**  
PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE

Rev. 5, 04/08/2011