



CERTIFICATE OF ANALYSIS

ChlorAC

1700-0132

Product: ChorAC

Part Number: 1700-0132

Lot Number: 107019

Production Date: 18 JUL 2011

Expiration Date: 14 JUL 2012

Primary Application: Preservative for waters samples in Post-column analysis of carbamate pesticides

Acceptance Criteria:

1. pH = 3.0 ± 0.2 @ $25 \text{ }^\circ\text{C} \pm 0.4 \text{ }^\circ\text{C}$ (Y/N) y
2. No interfering peaks in blank (Y/N) y
3. Twelve appropriate peaks in the chromatogram (Y/N) y

Sample: Carbamate Test Mixture, **PN:** 1700-0063

Storage Conditions: Dry at 8°C to 25°C

Analysis Certified by: *Jim Murphy* **Date:** 21 Jul 2011

VP OPER
Title

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE