



CERTIFICATE OF ANALYSIS

ChlorAC

1700-0132

Product: ChorAC
Part Number: 1700-0132
Lot Number: 112019
Production Date: 15 Dec 2011
Expiration Date: 14 Dec 2012

Primary Application: Preservative for waters samples in Post-column analysis of carbamate pesticides

Acceptance Criteria:

- 1. pH = 3.0 ± 0.2 @ 25 ± 0.4 °C (Y/N)
(for 3% solution of ChlorAc)
- 2. No interfering peaks in blank (Y/N)
- 3. Twelve appropriate peaks in chromatogram (Y/N)

Y
Y
Y

Sample: Carbamate Test Mixture, PN: 1700-0063

Storage Conditions: Dry at 8°C to 25°C

Analysis Certified by: *Jim Murphy* **Date:** 19 MAR 2012
VP OPER
Title

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

P:\Documentation\Chemical Documentation\CAc 1700-0132

Rev. A, 06/17/2011