



## CERTIFICATE OF ANALYSIS

**ChlorAC**

**1700-0132**

**Product:** ChorAC

**Part Number:** 1700-0132

**Lot Number:** 504014

**Production Date:** 08 Apr 2015

**Expiration Date:** 08 Apr 2016

**Primary Application:** Preservative for waters samples in Post-column analysis of carbamate pesticides

**Acceptance Criteria:**

- |   |          |
|---|----------|
| 1. pH = $3.0 \pm 0.2$ @ $25 \pm 0.4$ °C (Y/N)<br>(for 3% solution of ChlorAc) | <u>y</u> |
| 2. No interfering peaks in blank (Y/N)  | <u>y</u> |
| 3. Twelve appropriate peaks in chromatogram (Y/N)                             | <u>y</u> |

**Sample:** Carbamate Test Mixture, **PN:** 1700-0063

**Storage Conditions:** Dry at 8°C to 25°C

**Analysis Certified by:** Jim Murphy **Date:** April 13, 2015  
President  
Title