



## CERTIFICATE OF ANALYSIS

**ChlorAC**

**1700-0132**

**Product:** ChorAC

**Part Number:** 1700-0132

**Lot Number:** 506007

**Production Date:** 04 Jun 2015

**Expiration Date:** 04 Jnu 2016

**Primary Application:** Preservative for waters samples in Post-column analysis of carbamate pesticides

**Acceptance Criteria:**

1. pH =  $3.0 \pm 0.2$  @  $25 \pm 0.4$  °C (Y/N)  
(for 3% solution of ChlorAc) y
2. No interfering peaks in blank (Y/N) y
3. Twelve appropriate peaks in chromatogram (Y/N) y

**Sample:** Carbamate Test Mixture, **PN:** 1700-0063

**Storage Conditions:** Dry at 8°C to 25°C

**Analysis Certified by:** Jim Murphy **Date:** 08 Jun 2015  
President  
Title