



CERTIFICATE OF ANALYSIS

RESTORE

1700-0140

Product: RESTORE
Part Number: 1700-0140
Lot Number: 101002
Test Start Date: 04 JAN 2011

Primary Application: Removes metal ion contamination from ion-exchange guard or analytical columns in glyphosate analysis.

Acceptance Criteria:

Actual Values:

1. pH = 1.30 ± 0.05 @ 25 ± 0.4 °C - 1.254

Analysis Certified by: *Jim Murphy* Date: 11 FEB 2011

VP OPR
Title

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE