



Certificate of Analysis

Product Artificial Saliva ASTM E2720-16/ ASTM E2721-16 with Mucin, Stabilized
Part Number 1700-0316
Primary Application Product testing and research
Volume 200 mL
Lot Number 2108018
Production Date 10 August 2021
Expiration Date 10 August 2022
Storage Conditions Dry at 8° to 25°C
Acceptance Criteria

1. Appearance:.....PASS
2. pH = 7.0 ± 0.2 @ $25 \pm 0.4^\circ\text{C}$:.....PASS
3. Other components gravimetrically-certified:.....PASS

Components


Sodium Chloride	Potassium Chloride	Magnesium Chloride	Mucin
Ammonium Chloride	Potassium Thiocyanate	Sodium Bicarbonate	Preservative
Potassium Phosphate	Calcium Chloride	Urea	

*All product components are weighed on NIST-traceable scales and measured in Class A volumetric glassware

Analysis Certified By: Maria Ofitserova, Ph.D.

Date: 18 August 2021

The product is manufactured, packaged, stored and shipped according to manufacturing standard best practices and under modern sanitary conditions.



CERTIFIED