



## CERTIFICATE OF ANALYSIS

### Artificial Perspiration, ISO 11641, Stabilized 1700-0509

**Product:** Artificial Perspiration ISO 11641, Stabilized

**Part Number:** 1700-0509

**Lot Number:** 806033

**Production Date:** 22 Jun 2018

**Expiration Date:** 22 Jun 2020

**Primary Application:** Determination of color fastness to perspiration (leather)

**Acceptance Criteria:**

1. pH =  $8.0 \pm 0.1$  @  $25 \pm 0.4$  °C (Y/N)

y

**Storage Conditions:** Dry at 8 ° to 25° C

**Analysis Certified by:** Jim Murphy  
President

Title

**Date:** June 25, 2018

Rev. B, 10/22/15