



Certificate of Analysis

Product Artificial Perspiration, ISO 3160, Custom pH, Stabilized
Part Number 1700-0526
Primary Application Determination of corrosion resistance
Volume 200 mL
Lot Number 2110004
Production Date 05 October 2021
Expiration Date 05 October 2022
Storage Conditions Dry at 8° to 25°C
Acceptance Criteria

1. Appearance:.....PASS
2. pH = 4.5 ± 0.1 @ $25 \pm 0.4^\circ\text{C}$:.....PASS
3. Other components gravimetrically-certified:.....PASS

Components

DI Water	Acetic Acid	Preservative
Sodium Chloride	Ammonium Chloride	
DL-Lactic Acid	Urea	

*All product components are weighed on NIST-traceable scales and measured in Class A volumetric glassware

Analysis Certified By: Maria Ofitserova, Ph.D.

Date: 14 October 2021

The product is manufactured, packaged, stored and shipped according to manufacturing standard best practices and under modern sanitary conditions.



CERTIFIED