



CERTIFICATE OF ANALYSIS

Artificial Perspiration, AATCC (Stabilized) 1700-0541

Product: Artificial Perspiration according to AATCC Test Method 15 (Stabilized)

Part Number: 1700-0541

Lot Number: 008006

Production Date: August 4, 2020

Expiration Date: August 4, 2021

Primary Application: Determination of color fastness to perspiration

Acceptance Criteria:

1. pH = 4.3 ± 0.1 @ 25 ± 0.4 °C (Y/N) Y

Storage Conditions: Store dry at 8 ° to 25° C

Analysis Certified by: Rebecca Smith **Date:** August 12, 2020
Vice President
Title

Rev. A, 07/07/15