



CERTIFICATE OF ANALYSIS

Artificial Perspiration, AATCC (Stabilized) 1700-0555

Product: Artificial Perspiration according to AATCC Test Method 15 (Stabilized)

Part Number: 1700-0555

Lot Number: 912027

Production Date: 12 Dec 2019

Expiration Date: 12 Dec 2020

Primary Application: Determination of color fastness to perspiration

Acceptance Criteria:

1. pH = 4.3 ± 0.1 @ 25 ± 0.4 °C (Y/N)

y

Storage Conditions: Store dry at 8 ° to 25° C

Analysis Certified by: Jim Murphy

President

Title

Date: January 3, 2020

Rev. A, 04/12/2016