



CERTIFICATE OF ANALYSIS

Artificial Urine DIN EN 1616:1999, Stabilized 1700-0558

Product: Artificial Urine DIN EN 1616:1999, stabilized

Part Number: 1700-0558

Lot Number: 2104013

Production Date: April 7, 2021

Expiration Date: April 7, 2022

Primary Application: For testing of urethral catheters.

Acceptance Criteria:

1. pH = 6.6 ± 0.1 @ 25 ± 0.4 °C (Y/N)

Y

Storage Conditions: Dry at 8 ° to 25° C

Analysis Certified by: Jim Murphy

President

Title

Date: April 20, 2021

Rev. A, 05/14/19