



CERTIFICATE OF ANALYSIS

Artificial Perspiration
DIN 53160-2:200-10, BS EN 1811:2011
Buffered, Custom pH
1700-0569

Product: Artificial Perspiration according to DIN 53160-2:200-10, BS EN 1811:201, buffered, Custom pH

Part Number: 1700-0569

Lot Number: 2102009

Production Date: February 4, 2021

Expiration Date: February 4, 2022

Primary Application: Product testing

Acceptance Criteria:

pH = 6.0 ± 0.1 @ 25 ± 0.4 °C (Y/N) Y

Storage Conditions: Dry at 8 ° to 25° C

Analysis Certified by: Jim Murphy **Date:** March 1, 2021
President
Title

Rev. A, 06/26/20