



## CERTIFICATE OF ANALYSIS

### Artificial Cerumen 1700-0701

**Product:** Artificial Cerumen  
**Part Number:** 1700-0701  
**Lot Number:** 902001  
**Production Date:** 01 Feb 2019  
**Expiration Date:** 01 Feb 2020

**Primary Application:** To be used for product testing

#### Specifications:

Composition as listed (Y/N) y

Oleic Acid - 20% w/w  
Linoleic acid - 20% w/w  
Lanolin - 20% w/w  
Paraffin Oil - 10% w/w  
Palmitic Acid - 20% w/w  
Myristic Acid - 10% w/w

**Storage Conditions:** Store refrigerated under nitrogen

**Analysis Certified by:** Jim Murphy **Date:** February 4, 2019  
President  
**Title**

Rev. A, 01/04/16