



CERTIFICATE OF ANALYSIS

KO1 Alkion Eluant

1700-1101

Product: KO1 Alkion Eluant for Polyamines Analysis

Part Number: 1700-1101

Lot Number: 011006

Test Date: 03 Nov 2010

Expiration Date: 03 Nov 2012

Primary Application:

1. Eluant for ALKION Cation-exchange chromatography of polyamines.

Acceptance Criteria:

1. pH = 6.00 ± 0.05 @ 25 ± 0.4 °C

2. Conductivity = 9.7 ± 0.15 mS

Actual Values:

6.033

9.72

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: _____

VP Ops
Title

Date: _____

05 Nov 2010

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE