



CERTIFICATE OF ANALYSIS

KO1 ALKION Eluant

1700-1101

Product: KO1 ALKION Eluant

Part Number: 1700-1101

Lot Number: 110031

Production Date: 28 Oct 2011

Expiration Date: 28 Oct 2013

Primary Application:

1. Eluant for ALKION cation-exchange column.


Acceptance Criteria:

1. pH = 6.00 ± 0.05 @ 25 ± 0.4 °C (Y/N)

Y

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:


VP Oper
Title

Date:

08 Nov 2011

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE