



CERTIFICATE OF ANALYSIS

KO1 ALKION Eluant

1700-1101

Product: KO1 ALKION Eluant

Part Number: 1700-1101

Lot Number: 802012

Production Date: 07 Feb 2018

Expiration Date: 07 Feb 2020

Primary Application:

1. Eluant for ALKION cation-exchange column.

Acceptance Criteria:

1. pH = 6.00 ± 0.05 @ 25 ± 0.4 °C (Y/N)

Y

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: Jim Murphy
President
Title

Date: February 14, 2018