



CERTIFICATE OF ANALYSIS

KO1 ALKION Eluant

1700-1101

Product: KO1 ALKION Eluant

Part Number: 1700-1101

Lot Number: 809009

Production Date: 05 Sep 2018

Expiration Date: 05 Sep 2020

Primary Application:

1. Eluant for ALKION cation-exchange column.

Acceptance Criteria:

1. pH = 6.00 ± 0.05 @ 25 ± 0.4 °C (Y/N)

y

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: Jim Murphy
President

Title

Date: Sept. 12, 2018