



## CERTIFICATE OF ANALYSIS

**KO1 ALKION Eluant**

**1700-1101**

**Product:** KO1 ALKION Eluant

**Part Number:** 1700-1101

**Lot Number:** 906008

**Production Date:** June 6, 2019

**Expiration Date:** June 6, 2021

**Primary Application:**

1. Eluant for ALKION cation-exchange column.

**Acceptance Criteria:**

1. pH =  $6.00 \pm 0.05$  @  $25 \pm 0.4$  °C (Y/N)

Y

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:** Rebecca Smith  
Director of Operations  
Title

**Date:** 6/10/2019