



# CERTIFICATE OF ANALYSIS

## KO2 ALKION Eluant

### 1700-1102

Product: KO2 ALKION Eluant  
Part Number: 1700-1102  
Lot Number: 005022  
Test Date: 20 MAY 2010  
Expiration Date: 20 MAY 2012

#### Primary Application:

1. Used in aminoglycoside antibiotics analyses.

#### Acceptance Criteria:

1. pH =  $12.8 \pm 0.10$  @  $25 \pm 0.4$  °C
2. Conductivity =  $24.9 \pm 0.2$  mS @  $25 \pm 0.4$  °C

#### Actual Values:

12.790  
24.8

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: \_\_\_\_\_

*[Signature]*  
VP OPR  
Title

Date: 21 MAY 2010

**ACCEPTED**  
PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE