



CERTIFICATE OF ANALYSIS

KO3 ALKION Eluant

1700-1103

Product: KO3 ALKION Eluant

Part Number: 1700-1103

Lot Number: 110010

Production Date: 07 Oct 2011

Expiration Date: 07 Oct 2013

Primary Application:

1. Eluant for ALKION cation-exchange column.

Acceptance Criteria:

1. pH = 5.7 ± 0.6 @ 25 ± 0.4 °C (Y/N)

Y

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:

Jim P. [Signature]
VP Oper
Title

Date:

19 Oct 2011

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE