



CERTIFICATE OF ANALYSIS

KO3 ALKION Eluant

1700-1103

Product: KO3 ALKION Eluant

Part Number: 1700-1103

Lot Number: 309003

Production Date: 05 Sep 2013

Expiration Date: 05 Sep 2015

Primary Application:

1. Eluant for ALKION cation-exchange column.

Acceptance Criteria:

1. pH = 5.7 ± 0.6 @ 25 ± 0.4 °C (Y/N)

y

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: Jim Murphy
President
Title

Date: 13 Sep 2013