



CERTIFICATE OF ANALYSIS

KO3 ALKION Eluant

1700-1103

Product: KO3 ALKION Eluant

Part Number: 1700-1103

Lot Number: 906010

Production Date: June 7, 2019

Expiration Date: June 7, 2021

Primary Application:

1. Eluant for ALKION cation-exchange column.

Acceptance Criteria:

1. pH = 5.7 ± 0.6 @ 25 ± 0.4 °C (Y/N)

Y

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: Rebecca Smith
Director of Operations
Title

Date: 6/11/2019