



# CERTIFICATE OF ANALYSIS

## SPE Sample Clean-up Cartridges

1705-0001

**Product:** SPE Sample Clean-up Cartridges

**Part Number:** 1705-0001

**Lot Number:** 406035

**Production Date:** 30 Jun 2014

**Expiration Date:** 30 Jun 2015

**Primary Application:** For sample clean-up prior to Glyphosate analysis

### Acceptance Criteria:

1. Amount of resin per cartridge is 2 g (Y/N)

y

2. Recoveries of Glyphosate from  
standard solution 95-105% (Y/N)

y

**Storage Conditions:** Dry at 8°C to 25°C

**Analysis Certified by:** Jim Murphy

**Date:** 07 Jul 2014

President

Title