



# CERTIFICATE OF ANALYSIS

## SPE columns, Cation-Exchange

**1705-0001**

**Product:** SPE Sample Clean-up Cartridges

**Part Number:** 1705-0001

**Lot Number:** 804025

**Production Date:** 17 Apr 2018

**Expiration Date:** 17 Apr 2019

**Primary Application:** For sample clean-up prior to Glyphosate analysis

### Acceptance Criteria:

1. Amount of resin per cartridge is 2 g (Y/N)

y

2. Recoveries of Glyphosate from  
standard solution 95-105% (Y/N)

y

**Storage Conditions:** Dry at 8°C to 25°C

**Analysis Certified by:** Jim Murphy  
President

Title

**Date:** April 26, 2018