



## CERTIFICATE OF ANALYSIS

### SPE columns, Cation-Exchange

1705-0001

**Product:** SPE Sample Clean-up Cartridges

**Part Number:** 1705-0001

**Lot Number:** 902046

**Production Date:** 27 Feb 2019

**Expiration Date:** 27 Feb 2020

**Primary Application:** For sample clean-up prior to Glyphosate analysis

#### Acceptance Criteria:

1. Amount of resin per cartridge is 2 g (Y/N)

y

2. Recoveries of Glyphosate from  
standard solution 95-105% (Y/N)

y

**Storage Conditions:** Dry at 8°C to 25°C

**Analysis Certified by:** Jim Murphy  
President

Title

**Date:** March 7, 2019