



CERTIFICATE OF ANALYSIS

Hydrolysis Reagent for Post-column Carbamate Pesticides Analysis

CB130.2

Product: Hydrolysis Reagent for Post-column analysis of Carbamate Pesticides
Part Number: CB130.2
Lot Number: 103009
Production Date: 03 MAR 2011
Expiration Date: 03 MAR 2013

Primary Application: Post-column Carbamate Pesticides analysis.

Acceptance Criteria:

1. pH = 12.5 ± 0.1 @ 25 ± 0.4 °C (Y/N)
2. Baseline noise (0 ± 25 mvolts) (Y/N)
3. Twelve appropriate peaks in chromatogram (Y/N)

Sample: Carbamate Test Mixture, PN 1700-0063,

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: *[Signature]* Date: 14 MAR 2011
VP Ops
Title

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev. A, 04/08/2011

P:\Documentation\Chemical Documentation\CB130.2