



CERTIFICATE OF ANALYSIS

Hydrolysis Reagent for Post-column Carbamate Pesticides Analysis for EPA Method 531.2

CB130.2

Product: Hydrolysis Reagent for Post-column analysis of Carbamate Pesticides

Part Number: CB130.2

Lot Number: 202042

Production Date: 29 Feb 2012

Expiration Date: 01 MAR 2014

Primary Application: Post-column Carbamate Pesticides analysis.

Acceptance Criteria:


1. pH = 12.6 ± 0.1 @ 25 ± 0.4 °C (Y/N)
2. Baseline noise (0 ± 25 mvolts) (Y/N)
3. Twelve appropriate peaks in chromatogram (Y/N)

4
4
4

Sample: Carbamate Test Mixture, PN 1700-0063,

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:


Title

Date:

19 MAR 2012

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev. A, 09/09/2011

P:\Documentation\Chemical Documentation\CB130.2