



CERTIFICATE OF ANALYSIS

Hydrolysis Reagent for Post-column Carbamate Pesticides Analysis for EPA Method 531.2

CB130.2

Product: Hydrolysis Reagent for Post-column analysis of Carbamate Pesticides

Part Number: CB130.2

Lot Number: 403010

Production Date: 10 Mar 2014

Expiration Date: 10 Mar 2016

Primary Application: Post-column Carbamate Pesticides analysis.

Acceptance Criteria:

- | | |
|---|----------|
| 1. pH = 12.6 ± 0.1 @ 25 ± 0.4 °C (Y/N) | <u>y</u> |
| 2. Baseline noise (0 ± 25 mvolts) (Y/N) | <u>y</u> |
| 3. Twelve appropriate peaks in chromatogram (Y/N) | <u>y</u> |

Sample: Carbamate Test Mixture, **PN** 1700-0063,

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: Jim Murphy **Date:** 17 Mar 2014
President
Title