



## CERTIFICATE OF ANALYSIS

### Hydrolysis Reagent for Post-column Carbamate Pesticides Analysis for EPA Method 531.2

**CB130.2**

**Product:** Hydrolysis Reagent for Post-column analysis of Carbamate Pesticides

**Part Number:** CB130.2

**Lot Number:** 805018

**Production Date:** 09 May 2018

**Expiration Date:** 09 May 2020

**Primary Application:** Post-column Carbamate Pesticides analysis.

**Acceptance Criteria:**

- |   |          |
|---|----------|
| 1. pH = $12.6 \pm 0.1$ @ $25 \pm 0.4$ °C (Y/N)    | <u>y</u> |
| 2. Baseline noise ( $0 \pm 25$ mvolts) (Y/N)      | <u>y</u> |
| 3. Twelve appropriate peaks in chromatogram (Y/N) | <u>y</u> |

**Sample:** Carbamate Test Mixture, **PN** 1700-0063,

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:** Jim Murphy **Date:** May 14, 2018  
President  
Title