



## CERTIFICATE OF ANALYSIS

### Hydrolysis Reagent for Post-column Carbamate Pesticides Analysis

**CB130**

**Product:** Hydrolysis reagent for post-column analysis of Carbamate pesticides

**Part Number:** CB130

**Lot Number:** 704018

**Production Date:** 12 Apr 2017

**Expiration Date:** 12 Apr 2019

**Primary Application:** Post-column analysis of Carbamate pesticides.

#### Acceptance Criteria

- |   |          |
|---|----------|
| 1. pH = $12.5 \pm 0.1$ @ $25 \pm 0.4$ °C (Y/N)    | <u>Y</u> |
| 2. Baseline noise ( $0 \pm 25$ mvolts) (Y/N)      | <u>Y</u> |
| 3. Twelve appropriate peaks in chromatogram (Y/N) | <u>Y</u> |

**Sample:** Carbamate Test Mixture, **PN:** 1700-0063.

**Storage Conditions:** Dry at 8°C to 25 °C

**Analysis Certified by:** Tony McIsaac **Date:** 17 Apr 2017  
VP Manufacturing  
Title