



## CERTIFICATE OF ANALYSIS

### Carbamate Analysis Hydrolysis Reagent

CB130

Product: Carbamate Analysis Hydrolysis Reagent

Part Number: CB130

Lot Number: 005019

Test Start Date: 18 MAY 2010

Expiration Date: 17 MAY 2012


Primary Application: Carbamate analysis.

#### Acceptance Criteria:

1. pH =  $12.5 \pm 0.1$  @  $25 \pm 0.4$  °C 12.43
2. Baseline noise ( $0 \pm 25$  mvolts) (Y/N) Y
3. Twelve appropriate peaks in chromatogram (Y/N) Y

Calibration Standard: Carbamate Test Standard, PN 1700-0063, LN 908045

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:  Date: 21 MAY 2010  
VP Ops  
Title

**ACCEPTED**  
PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE

Rev. 7, 9/16/08