



CERTIFICATE OF ANALYSIS
Carbamate Analysis Hydrolysis Reagent
CB130

Product: Carbamate Analysis Hydrolysis Reagent
Part Number: CB130
Lot Number: 006005
Test Start Date: 04 JUN 2010
Expiration Date: 03 JUN 2012


Primary Application: Carbamate analysis.

Acceptance Criteria:

- 1. pH = 12.5 ± 0.1 @ 25 ± 0.4 °C 12.499
- 2. Baseline noise (0 ± 25 mvolts) (Y/N) Y
- 3. Twelve appropriate peaks in chromatogram (Y/N) Y

Calibration Standard: Carbamate Test Standard, PN 1700-0063, LN 908045

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:  **Date:** 10 Jun 2010
V.P. OPR.
Title

Rev. 7, 9/16/08

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE