



CERTIFICATE OF ANALYSIS

Carbamate Analysis Hydrolysis Reagent

CB130

Product: Carbamate Analysis Hydrolysis Reagent

Part Number: CB130

Lot Number: 007006

Test Start Date: 12 Jul 2010

Expiration Date: 09 Jul 2012

Primary Application: Carbamate analysis.

Acceptance Criteria:

1. pH = 12.5 ± 0.1 @ 25 ± 0.4 °C

12.459

2. Baseline noise (0 ± 25 mvolts) (Y/N)

Y

3. Twelve appropriate peaks in chromatogram (Y/N)

Y

Calibration Standard: Carbamate Test Standard, PN 1700-0063, LN 908045

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: 

Date: 15 Jul 2010

VP OPR

Title

Rev. 7, 9/16/08

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE