



# CERTIFICATE OF ANALYSIS

## Carbamate Analysis Hydrolysis Reagent

**CB130**

**Product:** Carbamate Analysis Hydrolysis Reagent

**Part Number:** CB130

**Lot Number:** 007024

**Test Start Date:** 20 Jul 2010

**Expiration Date:** 20 Jul 2012


**Primary Application:** Carbamate analysis.

### Acceptance Criteria:

1. pH =  $12.5 \pm 0.1$  @  $25 \pm 0.4$  °C 12.519
2. Baseline noise ( $0 \pm 25$  mvolts) (Y/N) Y
3. Twelve appropriate peaks in chromatogram (Y/N) Y

**Calibration Standard:** Carbamate Test Standard, PN 1700-0063, LN 908045

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:**  **Date:** 27 Jul 2010  
VP OPR  
Title

Rev. 7, 9/16/08

**ACCEPTED**  
PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE