

CERTIFICATE OF ANALYSIS
Carbamate Analysis Hydrolysis Reagent
CB130

Product: Carbamate Analysis Hydrolysis Reagent

Part Number: CB130

Lot Number: 009043

Test Start Date: 28 Sep 2010

Expiration Date: 27 Sep 2012

Primary Application: Carbamate analysis.

Acceptance Criteria:

1. pH = 12.5 ± 0.1 @ 25 ± 0.4 °C

12.536

2. Baseline noise (0 ± 25 mvolts) (Y/N)

Y

3. Twelve appropriate peaks in chromatogram (Y/N)

Y

Calibration Standard: Carbamate Test Standard, PN 1700-0063, LN

908045

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:

[Signature]
VP QPR

Title

Date:

30 Sep 2010