



# CERTIFICATE OF ANALYSIS

## Carbamate Analysis Hydrolysis Reagent

**CB130**

**Product:** Carbamate Analysis Hydrolysis Reagent

**Part Number:** CB130

**Lot Number:** 011001

**Test Start Date:** 02 Nov 2010

**Expiration Date:** 01 Nov 2012

**Primary Application:** Carbamate analysis.

**Acceptance Criteria:**

- 1. pH =  $12.5 \pm 0.1$  @  $25 \pm 0.4$  °C 12.489
- 2. Baseline noise ( $0 \pm 25$  mvolts) (Y/N) Y
- 3. Twelve appropriate peaks in chromatogram (Y/N) Y

**Calibration Standard:** Carbamate Test Standard, PN 1700-0063, LN 908045

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:** *Jim Murphy* **Date:** 08 Nov 2010  
VPOK  
Title

**ACCEPTED**  
PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE

Rev. 7, 9/16/08