



## CERTIFICATE OF ANALYSIS

### Hydrolysis Reagent for Post-column Carbamate Pesticides Analysis

**CB130**

**Product:** Hydrolysis Reagent for post-column analysis of Carbamate pesticides

**Part Number:** CB130

**Lot Number:** 102034

**Expiration Date:** 23 Feb 2013

**Test Start Date:** 23 Feb 2011

**Primary Application:** Post-column analysis of Carbamate pesticides.

**Acceptance Criteria:**

1. pH =  $12.5 \pm 0.1$  @  $25 \pm 0.4$  °C
2. Baseline noise ( $0 \pm 25$  mvolts) (Y/N)
3. Twelve appropriate peaks in chromatogram (Y/N)

12-578  
Y  
Y

**Sample:** Carbamate Test Mixture, **PN:** 1700-0063, **LN:** 908045

**Storage Conditions:** Dry at 8°C to 25 °C

**Analysis Certified by:**

**Title:**

VP OPR

**Date:** 02 MAR 2011

Rev. 8, 02/15/11

**ACCEPTED**  
PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE