



## CERTIFICATE OF ANALYSIS

### Hydrolysis Reagent for Post-column Carbamate Pesticides Analysis

**CB130**

**Product:** Hydrolysis reagent for post-column analysis of Carbamate pesticides

**Part Number:** CB130

**Lot Number:** 108028

**Production Date:** 25 AUG 2011

**Expiration Date:** 23 AUG 2013

**Primary Application:** Post-column analysis of Carbamate pesticides.

#### Acceptance Criteria

1. pH =  $12.5 \pm 0.1$  @  $25 \pm 0.4$  °C (Y/N)
2. Baseline noise ( $0 \pm 25$  mvolts) (Y/N)
3. Twelve appropriate peaks in chromatogram (Y/N)

Y  
Y  
Y

**Sample:** Carbamate Test Mixture, PN: 1700-0063.

**Storage Conditions:** Dry at 8°C to 25 °C

**Analysis Certified by:** \_\_\_\_\_

*Jim Murphy*  
VP OPR  
Title

**Date:** 20 Sep 2011

# ACCEPTED

PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE

Rev. A, 04/08/2011

P:\Documentation\Chemical Documentation\CB130