



## CERTIFICATE OF ANALYSIS

### Hydrolysis Reagent for Post-column Carbamate Pesticides Analysis

CB130

Product: Hydrolysis reagent for post-column analysis of Carbamate pesticides

Part Number: CB130

Lot Number: 202002

Production Date: 01 Feb 2012

Expiration Date: 01 Feb 2014

Primary Application: Post-column analysis of Carbamate pesticides.

#### Acceptance Criteria

1. pH =  $12.5 \pm 0.1$  @  $25 \pm 0.4$  °C (Y/N)
2. Baseline noise ( $0 \pm 25$  mvolts) (Y/N)
3. Twelve appropriate peaks in chromatogram (Y/N)

Y  
Y  
Y

Sample: Carbamate Test Mixture, PN: 1700-0063.

Storage Conditions: Dry at 8°C to 25 °C

Analysis Certified by:

*Jim Murphy*  
*J.P. O'Neil*  
Title

Date:

28 Feb 2012

# ACCEPTED

PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE