



## CERTIFICATE OF ANALYSIS

### Hydrolysis Reagent for Post-column Carbamate Pesticides Analysis

**CB130**

**Product:** Hydrolysis reagent for post-column analysis of Carbamate pesticides

**Part Number:** CB130

**Lot Number:** 2030 20

**Production Date:** 16 MAR 2012

**Expiration Date:** 16 MAR 2014

**Primary Application:** Post-column analysis of Carbamate pesticides.

#### Acceptance Criteria

1. pH =  $12.5 \pm 0.1$  @  $25 \pm 0.4$  °C (Y/N)
2. Baseline noise ( $0 \pm 25$  mvolts) (Y/N)
3. Twelve appropriate peaks in chromatogram (Y/N)

4  
4  
4

**Sample:** Carbamate Test Mixture, PN: 1700-0063.

**Storage Conditions:** Dry at 8°C to 25 °C

**Analysis Certified by:**

*Jim Young*  
V.P. Oper  
Title

**Date:**

30 MAR 2012

# ACCEPTED

PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE