



## CERTIFICATE OF ANALYSIS

### OPA Diluent for Post-column Carbamate Pesticides Analysis

#### CB910

**Product:** OPA Diluent for Post-column analysis of Carbamate pesticides

**Part Number:** CB910

**Lot Number:** 508022

**Production Date:** 13 Aug 2015

**Expiration Date:** 13 Aug 2017

**Primary Application:** Post-column analysis of Carbamate pesticides.

#### Acceptance Criteria

- |   |          |
|---|----------|
| 1. pH = $9.10 \pm 0.1$ @ $25 \pm 0.4$ °C (Y/N)    | <u>y</u> |
| 2. Baseline noise ( $0 \pm 25$ mvolts) (Y/N)      | <u>y</u> |
| 3. Twelve appropriate peaks in chromatogram (Y/N) | <u>y</u> |

**Sample:** Carbamate Test Mixture, **PN:** 1700-0063

**Storage Conditions:** Dry at 8 °C to 25 °C

**Analysis Certified by:** Jim Murphy **Date:** August 17, 2015  
President  
Title