



# CERTIFICATE OF ANALYSIS

## OPA Diluent for Post-Column Carbamate Analysis

### CB910

Product: OPA Diluent for Post Column Carbamate Analysis

Part Number: CB910

Lot Number: 008045

Test Start Date: 30 Sep 2010

Expiration Date: 30 Sep 2012

**Primary Application:** For preparation of post-column fluorescence reagent for carbamate pesticide analysis.

#### Acceptance Criteria:

1. pH =  $9.10 \pm 0.1$  @  $25 \pm 0.4$  °C 9.107
2. Baseline noise ( $0 \pm 25$  mvolts) (Y/N) Y
3. Twelve appropriate peaks in chromatogram (Y/N) Y

**Calibration Standard:** Carbamate Test Standard, PN 1700-0063, LN 908045

**Storage Conditions:** Dry at 8 to 25 °C

Analysis Certified by: *Jim Murphy* Date: 04 Oct 2010  
*J.P. Ope*  
Title

Rev. 8, 9/19/08

**ACCEPTED**  
PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE