



# CERTIFICATE OF ANALYSIS

## OPA Diluent for Post-Column Carbamate Analysis

CB910

Product: OPA Diluent for Post Column Carbamate Analysis

Part Number: CB910

Lot Number: 010097

Test Start Date: 29 Oct 2010

Expiration Date: 27 Oct 2012


**Primary Application:** For preparation of post-column fluorescence reagent for carbamate pesticide analysis.

### Acceptance Criteria:

1. pH =  $9.10 \pm 0.1$  @  $25 \pm 0.4$  °C 9.115
2. Baseline noise ( $0 \pm 25$  mvolts) (Y/N) Y
3. Twelve appropriate peaks in chromatogram (Y/N) Y

Calibration Standard: Carbamate Test Standard, PN 1700-0063, LN 908045

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:  Date: 02 Nov 2010  
VP OPR  
Title

**ACCEPTED**  
PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE

Rev. 8, 9/19/08