



CERTIFICATE OF ANALYSIS

OPA Diluent for Post-Column Carbamate Analysis

CB910

Product: OPA Diluent for Post Column Carbamate Analysis

Part Number: CB910

Lot Number: 011011

Test Start Date: 09 Nov 2010

Expiration Date: 08 Nov 2012

Primary Application: For preparation of post-column fluorescence reagent for carbamate pesticide analysis.

Acceptance Criteria:

1. pH = 9.10 ± 0.1 @ 25 ± 0.4 °C 9.18
2. Baseline noise (0 ± 25 mvolts) (Y/N) Y
3. Twelve appropriate peaks in chromatogram (Y/N) Y

Calibration Standard: Carbamate Test Standard, PN 1700-0063, LN 908045

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: *Jim Murphy* Date: 22 Nov 2010

VP Ops
Title

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE